Calibration of Geophone Microseismic Sensors

Gordon B. Bowden

Stanford Linear Accelerator Center, Stanford University, Stanford, CA 94309

Abstract

This note describes a simple lab bench technique for calibrating the sensitivity and dynamics of the Mark Products L4C geophones used at SLAC for measuring microseismic ground motion and mechanical vibrations of accelerator support structures. Calibration results for SLAC's three L4C geophones are recorded.

1 L4C geophone

The L4C geophone is a seismic instrument producing a voltage proportional to the velocity difference ($\dot{x} - \dot{x}_g$) between its proof mass M and its housing. Voltage is generated in a pickup coil wound on the proof mass when a permanent magnet attached to the housing moves past it. The suspension system consists of a 1 kilogram proof mass suspended on soft springs with resonant frequency around 1 Hz. A photo of the instrument and the instrument's schematic are shown below:

Figure 1: Vertical L4C geophone

In the righthand schematic, the proof mass M (kg) is suspended by spring stiffness k (n/m) and mechanically damped by dashpot b (n/m/sec). The motion sensing coil is mounted above proof mass M and is represented by inductance L_c in series with coil resistance R_c . The dynamics of this instrument are derived in NLC ME Note 1 [?]. In short it acts as a damped mechanical oscillator electromagnetically coupled to the output by the moving field of a permanent magnet attached to the housing. For ground motions above mechanical resonance, the output signal follows the ground velocity $\dot{x}_g$ ($\mu\text{m}/\text{sec}$). For frequencies below resonance, voltage falls as proof mass motion starts to follow ground motion. By its self, the L4C is underdamped (damping ratio $\zeta = .2$) and resonates with ground motion around 1 Hz delivering exaggerated output and shifting phase for seismic frequencies near its resonance. To damp this resonance it is customary to shunt the geophone output with an external resistor R_s sufficient to critically damp ($\zeta = .707$) the resonance. This damping resistor is connected across the output inside a small Lemo box on the output cable shown in the left photo of figure 1.

2 Calibration Procedure

Calibration is based on the step-release method described by Rodgers et al [?] and sketched in figure ???. In this method a DC current is applied to the geophone's pickup coil. This current displaces the instrument's proof mass from it equilibrium position. A double pole double throw switch is used to abruptly disconnect the geophone coil from the current source and connect it to a digitally recording oscilloscope which records the voltage output as the proof mass returns to equilibrium. The measured step release voltage transient is then numerically fit to the damped sinusoidal response of the geophone:

$$e(t) = \frac{K}{\omega} e^{-\sigma(t-t_o)} \sin(\omega(t-t_o)) u(t-t_o) \quad \text{volts} \quad (1)$$

In this equation $u(t-t_o)$ is a unit step function starting at time $t = t_o$ which accommodates the offset between the scope trigger and the start of motion. The other 3 constants to be fit which determine the geophone's dynamics are:

- $\omega \equiv$ geophone damped oscillation frequency (radians/sec)
- $\sigma \equiv$ geophone damping constant (1/sec)
- $K \equiv$ geophone step response constant (volts/sec)

This method of calibration is entirely electrical and does not involve shake tables or measuring micron displacements. Assuming a linear system, the only parameter which must be taken on faith from the manufacturer is the proof mass, an easily measured and stable instrument parameter.

Figure 2: bench top calibration of L4C geophone

3 Geophone Generator Constant

The main goal of the calibration is to determine the geophone's generator constant G (volts/(m/sec)) or (newtons/amp) which relates voltage to instrument velocity. G depends on the strength of the magnetic field and the design of the instrument's coil. Note that the same constant G determines both voltage and force in this electromechanical system:

- $v_{coil} = G(\dot{x} - \dot{x}_g)$ volts
- $F_{coil} = -Gi_{coil}$ newtons

During calibration it is assumed that the ground motion x_g is completely negligible compared to the coil motion x set up by the calibration current. The relation between G and output voltage waveform after release depends on initial conditions and the geophone's equations of motion. Before being switched from the power supply to the scope, the current in the geophone coil is $i_o = v_{in}/R_c$ where v_{in} is the measured supply voltage and R_c is the measured geophone coil resistance. This current is the same irrespective of any shunting damping resistor R_s across the geophone output. Current i_o causes proof mass displacement $x_o = -Gi_o/k = -Gv_{in}/R_ck$ where k is the spring stiffness (newtons/m) of the geophone's proof mass suspension. k is related to the proof mass and natural frequency by $k = M\omega_o^2$. A typical calibration displacement of the proof mass can be calculated using data from the calibration of geophone No.635:

$$x_o = \frac{Gv_{in}}{R_cM\omega_o^2} = \frac{(301 \text{ newtons/amp})(.954 \text{ volts})}{(5510 \Omega)(.958 \text{ newtons/m/sec}^2)(7.408 \text{ radian/sec})^2} = 0.991 \text{ mm} \quad (2)$$

When the DPDT switch disconnects the power supply the current stops abruptly. With no damping resistance shunting the output the inductive energy in the geophone coil is dissipated in a tiny arc at the switch contact, probably in microseconds. With a shunting resistor connected the time constant $L_c/(R_s + R_c) \simeq (6.05 \text{ henries})/(12 \text{ k}\Omega) \simeq .5 \text{ milli seconds}$. This is still a very brief interval compared to the 1 second oscillation period of the instrument. So the initial conditions upon release are $i = 0$, $x = x_o$, $\dot{x} = 0$. Ignoring $L_c \frac{di}{dt}$, the equations of motion are:

$$\begin{aligned} \text{proof mass force equilibrium} \quad M\ddot{x} + b\dot{x} + kx + Gi &= 0 \\ \text{voltage sum around circuit} \quad G\dot{x} - (R_s + R_c)i &= 0 \end{aligned} \quad (3)$$

The Laplace transforms of these 2 equations including initial condition x_o are

$$\begin{aligned} \text{force} \quad (Ms^2 + bs + k)X(s) - (Ms + b)x_o + GI(s) &= 0 \\ \text{volts} \quad GsX(s) - Gx_o - (R_c + R_s)I(s) &= 0 \end{aligned} \quad (4)$$

$X(s)$ can be eliminated between these 2 equations and a transfer function written between $I(s)$ and initial condition x_o . Substituting $I(s) = V_{sig}(s)/R_s$ which is true if the scope input impedance is high enough to ignore and substituting initial condition $x_o = Gv_{in}/R_ck$:

$$V_{sig}(s) = \left(\frac{R_s}{R_s + R_c} \right) \left(\frac{-G^2 v_{in}}{R_c M} \right) \left[\frac{1}{s^2 + 2\sigma s + \omega_o^2} \right] \quad (5)$$

$$\omega_o \equiv \sqrt{\frac{k}{M}}, \quad 2\sigma = \frac{1}{M} \left(b + \frac{G^2}{R_c + R_s} \right), \quad \omega \equiv \sqrt{\omega_o^2 - \sigma^2}, \quad \zeta \equiv \frac{\sigma}{\omega_o}$$

The inverse Laplace transform of eq(??) has the same form as eq(1):

$$v_{sig}(t) = \left(\frac{R_s}{R_s + R_c} \right) \left(\frac{-G^2 v_{in}}{R_c M} \right) \left[\frac{1}{\omega} e^{-\sigma t} \sin \omega t \right] \quad (6)$$

The step response depends on G^2 because G determines both the initial displacement and later the voltage output. Comparing derived eq(??) to the measured waveform $e(t) = \frac{K}{\omega} e^{-\sigma t} \sin \omega t$, eq(??), the geophone generator constant G can be extracted from a fit for K in the measured waveform using the manufacturer's value for M , measured resistances R_s & R_c and the voltage v_{in} used to initially displace the coil.

$$G = \sqrt{\left(\frac{R_s + R_c}{R_s} \right) \frac{MK}{v_{in}/R_c}} \quad (7)$$

G is the generator constant of the instrument by itself. Signals measured with a damping resistor R_s shunted across the output are attenuated so that the effective generator constant for a damped geophone is:

$$G_d = \left(\frac{R_s}{R_s + R_c} \right) G \quad (8)$$

4 Measurement Results

The procedure followed in calibrating the L4C geophones was to first measure the response directly without damping resistor. Then a resistor was selected which brought the step response to approximately critical damping. Response signals were digitized on a TEKTRONIX 3032 digital scope. Each response record is 10000 samples long. To simultaneously determine the best fit to the 4 parameters t_o , ω , σ and K in equation (??), a Matlab function **fminsearch** was used. To converge, this function requires a vector of initial guesses for $[t_o \ \omega \ \sigma \ K]$. Calibration results follow in Table 1. The first 2 geophones No.634 horiz. and No.635 vertical have quite similar responses but geophone No.636 has a natural frequency of 1.5 Hz instead of 1 Hz. Its damping is lower and it puts out a weaker signal. These 3 geophones were purchased in 1995. Images of their factory calibration data sheets are appended. All 3 factory calibration sheets look similar. Somehow in the intervening years, something has happened to No.636 to lower its damping and raise its resonant frequency. Possibly its mechanical suspension has been damaged. Figure(??) shows the response of all three geophones both without and with damping resistor R_s .

L4C Geophone Calibration Constants				
Serial No.	634 Horiz.	635 Vertical	636 Vertical	units
mass M	.9666	.9583	.9653	kg
coil res.(meas.) R_c	5430	5510	5470	Ω
No external damping resistance				
V_{step}	1.052	0.998	0.891	volts
f	1.0874	1.1462	1.5038	Hz
σ	1.6172	1.73478	1.7798	1/sec
ζ	0.2303	0.2342	0.1851	damping ratio
f_o	1.1174	1.1790	1.5302	Hz
K	17.164	17.119	13.244	volts/sec
G	292.64	300.95	280.15	volts/m/sec
G (mfg, 1995)	284.64	288.97	287.01	volts/m/sec
With external damping resistance				
R_s damping	6802	6487	3488	Ω
V_{step}	2.944	2.961	3.074	volts
f	0.6791	0.6276	0.8532	Hz
σ	5.335	4.9399	5.155	1/sec
ζ	0.7809	0.7815	0.6931	damping ratio
K	24.560	22.298	17.191	volts/sec
Damped Generator Constant				
G_d	162.73	162.73	109.08	volts/m/sec

Table 1: Geophone Calibration Results

Figure 3: Geophone Calibration Response

References

- [1] NLC ME Note No. 1-94 Rev 2, 'Mark L4C Geophone Design Constants', Gordon Bowden, 31 May 1995.
- [2] Rodgers, P.W., Martin, A.J., Robertson, M.C., Hsu, M.M. and Harris, D.B. (1995). Signal-Coil Calibration of Electromagnetic Seismometers, *Bull. Seism. Soc. Am.* **85**, no.3, 845-850.

INSPECTION DATA

L-4 C Horizontal

1. General

Serial Number: 634

- a) Case height: 13 cm.
- b) Case diameter: 7.6 cm.
- c) Total weight: 2.15 kg.
- d) Operating pressure under water: 500 PSI
- e) Polarity of voltages produced at A and C terminal when each suspended mass moves toward the bottom: Negative

2. Calibration Coil

- a) Turns: 12 Turns
- b) Resistance: 7.2 Ohms @ 68°F

3. Signal Coil

- a) Turns: 4950 turns on each of 2 coils connected in series/
- b) Electrodynamical constant: 7.23 V/In./Sec. (284.64 V/m/sec)
- c) Resistance: 5436 Ohms @ 68°F
- d) Leakage to case: > 100 Megohms at 500 volts
- e) Calibration constant: 46.7K Dynes/Ampere
- f) Frequency (fo) 100 HZ
- g) Suspended mass (m) 966.6 grams
- h) Open circuit damping (bo) .259 of critical damping

DATE: July, 26-1995

INSPECTOR: Lola Lenizy

INSPECTION DATA

L-4 C Vertical

1. General

Serial Number: 635

- a) Case height: 13 cm.
- b) Case diameter: 7.6 cm.
- c) Total weight: 2.15 kg.
- d) Operating pressure under water: 500 PSI
- e) Polarity of voltage produced at A and C terminal when each suspended mass moves toward the bottom Negative

2. Calibration Coil

- a) Turns: 12 turns.
- b) Resistance: 70 Ohms @ 68°F

3. Signal Coil

- a) Turns: 4950 turns on each of 2 coils connected in series/
- b) Electrodynamic constant: 7.34 V/In/Sec. (288.97 V/m/sec)
- c) Resistance: 5530 Ohms @ 68°F
- d) Leakage to case: >100 Megohms at 500 volts
- e) Calibration constant: 44.4 K Dynes/Ampere
- f) Frequency (fo) .990 Hz
- g) Suspended mass (m) 958.3 grams
- h) Open circuit damping (bo) .271 of critical damping

DATE: August, 03-1995

INSPECTOR: Lola Linzy

INSPECTION DATA

L-4 C Vertical

1. General

Serial Number: 636

- a) Case height: 13 cm.
- b) Case diameter: 7.6 cm.
- c) Total weight: 2.15 kg.
- d) Operating pressure under water: 500 PSI
- e) Polarity of voltage produced at A and C terminal when each suspended mass moves toward the bottom Negative

2. Calibration Coil

- a) Turns: 12 turns.
- b) Resistance: 7.1 Ohms @ 68°F

3. Signal Coil

- a) Turns: 4950 turns on each of 2 coils connected in series/
- b) Electrodynamic constant: 7.29 V/In/Sec. (287.01 V/m/Sec)
- c) Resistance: 5499 Ohms @ 68°F
- d) Leakage to case: >100 Megohms at 500 volts
- e) Calibration constant: 45.8 K Dynes/Ampere
- f) Frequency (fo) .990 Hz
- g) Suspended mass (m) 965.3 grams
- h) Open circuit damping (bo) .287 of critical damping

DATE: August, 03 - 1995 INSPECTOR: Lara Loring